



### CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 04.10.2023
		Calibration Valid: 03.04.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, GOVT.PRIMARY HEALTH CENTRE, THIRUMANUR, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	BDI-152
iii	Serial No	GPHC/TMR/CEN001
iv	Location	LABORATORY

### CALIBRATION WORKSHEET

#### Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	1999	-1

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,  
CBE – 04, hbmedicalsystem@gmail.com

If + Ve.  
Prog.  
Case No.

15