

**CALIBRATION REPORT**

|   |                 |  |
|---|-----------------|--|
| Report No: HBMS/CAL/23-24                           |                 | Calibration Date: 04.10.2023                           |
|   |                 | Calibration Valid: 03.03.2024                          |
| <b>I.1. CUSTOMER DETAILS:</b>                       |                 |  |
| Name and Address of the organization                |                 | M/s, GOVT.PRIMARY HEALTH CENTRE,<br>THIRUMANUR, SALEM. |
| <b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b> |                 |  |
| EQUIPMENT NAME                                      |                 | PIPETTE  |
| I   | Manufactured by | MICROLUX   |
| II  | Model           | 10µl   |
| III   | Serial No       | GPHC/TMR/PIP/001                                       |
| IV  | Location        | LABORATORY   |

**Calibration standards used :**

| Master instrument        | Model   | Serial no  | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022         | 06-OCT-2023        |

**Details of the machine:** machine calibrated and checked output.

| Number of measurement | Channel 1 (Volume) |
|-----------------------|--------------------|
|                       | 10µl               |
| 1                     | 10.05              |
| 2                     | 10.06              |
| 3                     | 10.01              |
| 4                     | 9.97               |
| 5                     | 10.01              |
| 6                     | 10.02              |



For Heart Beat Medical Systems  
(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016  
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106  
88833 15662

No.42/1, Navaindia Road, Peelamedu,  
CBE - 04, hbmedicalsystem@gmail.com