



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 04.10.2023
	Calibration Valid: 03.03.2024
I.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVT.PRIMARY HEALTH CENTRE, THIRUMANUR, SALEM.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I Manufactured by	MICROFILL
II Model	50µl
III Serial No	GPHC/TMR/PIP/002
IV Location	LABORATORY

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	50µl
1	50.03
2	50.09
3	50.05
4	49.96
5	50.01
6	50.08



For Heart Beat Medical Systems
(Authorized signatory)

[Handwritten Signature]

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,
CBE - 04, hbmedicalsystem@gmail.com