



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name: Govt. PHC, Veeraganur      Lab No: 12620  
 Constituent Group: Chemistry III      Date of Result Entered: 20/09/2023  
 PT Item: Lyophilized human serum based      Date of Report Published: 05/10/2023

Sl. No.	Analyte	Method / Principle	Instrument	No. of Measurements	CV	CV <sub>max</sub>	CV <sub>min</sub>	CV <sub>max</sub> - CV <sub>min</sub>	CV <sub>max</sub> / CV <sub>min</sub>	CV <sub>max</sub> - CV <sub>min</sub>
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	306.5 mg/dL	-0.29	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.35 mg/dL	0.57	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	117.9 mg/dL	0.99	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 Govt. PHC, Veeraganur  
 VEERAGANUR  
 Veeraganur  
 TAMILNADU636116

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*