



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 15.09.2023
		Calibration Valid: 14.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, GOVT.PRIMARY HEALTH CENTRE, ACHANKUTTAPATTY, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	BD INSTRUMENTATION
ii	Model	BDI-152
iii	Serial No	PHC/APY/CF/001
iv	Location	LABORATORY

CALIBRATION WORKSHEET

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	0
2	1200	1200	0
3	1500	1501	1
4	2000	1999	-1

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com