

**CALIBRATION REPORT**

Report No: HBMS/CAL/23-24		Calibration Date: 04.10.2023
		Calibration Valid: 03.03.2024
<b>I.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, GOVT.PRIMARY HEALTH CENTRE, THIRUMANUR, SALEM.
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	MICROLUX
II	Model	500µl
III	Serial No	GPHC/TMR/PIP/003
IV	Location	LABORATORY

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	500µl
1	500.02
2	500.08
3	500.04
4	499.96
5	500.05
6	500.09



**For Heart Beat Medical Systems**  
(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016  
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106  
88833 15662

No.42/1, Navaindia Road, Peelamedu,  
CBE - 04, hbmedicalsystem@gmail.com