



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2023



Lab Name: Govt. PHC, Vellalapatti
Constituent Group: Chemistry III
PT Item: Lyophilized Serum

Lab No: 12602
Date of Result Entered: 19/01/2023
Date of Report Published: 09/02/2023

Ch
10
C

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participant CV	SD	Total Value	SD	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	110.87	5.38	7.05	124 mg/dL	0.54	0.35
2	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	1.9 mg/dL	-1.40	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	5.03	6.76	120 mg/dL	1.15	0.32

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Vellalapatti
Vellalapatti primary health center
Vellalapatti
TAMILNADU 635012

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****