

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13614

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
ERASAKKANAYAKANUR-625515, THENI DISTRICT.

SRF No.	SRF/23/00637-0004
SRF Date	07-10-2023
Date of Receipt	07-10-2023
Date of Calibration	07-10-2023
Due Date for Calibration	06-10-2024
Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	10µl	Material	PVC
Serial Number	RW12253	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.8	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl


Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.87	99.87	99.85	99.84	99.89	99.87	-0.13	0.02	0.47
		99.89	99.84	99.90	99.87	99.90				
2	500	499.90	499.89	499.90	499.87	499.88	499.88	-0.12	0.02	0.47
		499.89	499.88	499.85	499.87	499.90				
3	1000	999.89	999.88	999.89	999.87	999.90	999.89	-0.11	0.03	0.47
		999.84	999.90	999.91	999.94	999.86				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13615 Page: No: 1 of 1

Customer Name & Address		SRF No.	
GOVERNMENT PRIMARY HEALTH CENTRE, ERASAKKANAYAKANUR-625515, THENI DISTRICT.		SRF/23/00637-0005	
		SRF Date	
		07-10-2023	
		Date of Receipt	
		07-10-2023	
		Date of Calibration	
		07-10-2023	
		Due Date for Calibration	
		06-10-2024	
		Issue Date	
		07-10-2023	

Details of Unit Under Calibration			
Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW09312	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.8	Humidity (%RH)	57	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used						
SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10	9.94	9.97	9.95	9.96	9.93	9.95	-0.05	0.02	0.47	
		9.98	9.95	9.92	9.97	9.96					
2	50	49.85	49.86	49.87	49.82	49.83	49.85	-0.15	0.02	0.47	
		49.87	49.85	49.82	49.89	49.87					
3	100	99.85	99.87	99.85	99.84	99.89	99.87	-0.13	0.02	0.47	
		99.89	99.87	99.89	99.87	99.90					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

M. Madhiban
(Calibration Engineer)
M.ADHIBAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.