




## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12485		MEDICAL DEVICES		Page No:1 of 1
Issue Date	30-09-2023			
SRF No & Date	BRF/23/00550-0001 & 30-09-2023			
Receipt Date	30-09-2023			
Calibration Date	30-09-2023			
Calibration Due	29-09-2024			
Customer Name & Address				
GOVERNMENT UPGRADED PRIMARY HEALTH CENTRE, KLIYANUR-604102,VILLUPURAM DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	ELECTRICAL SAFETY(MICROSCOPE)	Make & Model	LABOMED & VISION 2000	
Range	MULTI	Sr. No	130661088	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.5°C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY			
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 93	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 145	Uncertainty in % (±) 11.1
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 221	Uncertainty in % (±) 8.7

- REMARKS**
- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
  - The user should determine the suitability of the instrument for its intended use.
  - The recalibration interval should be determined on the user requirement.
  - The results stated in this certificate relate only to the item calibrated.
  - The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
  - Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By  
  
 (Calibration Engineer)  
 M.DINESH

Authorised Signatory

<input checked="" type="checkbox"/> Technical Manager (C.SHANMUGARAJ)	<input type="checkbox"/> Chief Executive
--	--



**SUNSHINE BIOMEDICAL SOLUTIONS**  
 No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.