

NAVEEN TECHNOLOGIES



No.389, Dharsha Garden, Kandappa Colony, Annadanapatty, Salem - 2.
Cell : 94880 30384, 87787 93052 Email : mallnaveentech@gmail.com Web : www.naveentechnologies.in

CALIBRATION REPORT	
Report No : NT/CAL/23-24/VGR/24	Calibration Date : 28.08.2023
	Calibration Valid: 27.02.2024
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, VEERGANUR, ATTUR, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	CENTRIFUGE
i	Manufactured by LABOMED
ii	Model -----
iii	Serial No PHC/VGR/CEN/001

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test in Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	2	1.12
2	1200	1201	2	1.23
3	1500	1502	4	1.56
4	2000	1999	1	1.175

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

