



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MARCH 2013



Lab Name: Govt. PHC, Vellalapati
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: CMC02
 Date of Result Entered: 28/03/2013
 Date of Report Published: 28/03/2013

Sl No	Analyte	Method/Equipment Name	Analyzer Name	No of Participants	CV	CV	CV	CV	CV	CV
1	GLUCOSE III	GOD-PGD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	87 mg/dL	4.88	8.24
2	CREATININE III	Jaffe's Kinetic - Absaure Pterate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	8.88	8.12	1.3 mg/dL	4.25	8.81
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	118.83	5.23	5.88	187 mg/dL	4.88	8.28
SDI Range										
Within -1.00 to +1.00										
Within ±1.01 to ±2.00										
Within ±2.01 to ±2.99										
Beyond ±3.0										
Interpretation										
Excellent.										
Good.										
Accept with caution. Warning Signal.										
Unacceptable performance. Action Signal.										

LAB ADDRESS :
 Govt. PHC, Vellalapati
 Vellalapati primary health center
 Vellalapati
 TAMILNADUE6012

Coordinator/ Contact Details:
 Email: clintqc@cmcvellore.ac.in
 Contact Number: 0418-2583105

Rande Chittabur
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Rande

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - FEBRUARY 2023



Lab Name: Govt. PHC, Vellalapatti
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 12623
Date of Result Entered: 26/02/2023
Date of Report Published: 08/03/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participates		Total Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	155.11	6.56	10.18	155 mg/dL	-0.01	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1872	3.40	9.00	0.31	3 mg/dL	-1.31	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1788	112.79	5.23	5.90	118 mg/dL	0.54	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Vellalapatti
Vellalapatti primary health center
Vellalapatti
TAMILNADU 636012

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Vellalapati
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 12802
 Date of Result Entered: 18/07/2023
 Date of Report Published: 07/08/2023

SL No	Analyte	Method / Principle Name	Analyzer Name	Total Participants	CV	CV	CV	CV	CV	CV
1	GLUCOSE III	GOOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.97	106 mg/dL	-0.25	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Phosphate III	Robonik (Semi automatic / Fully Automatic)	1890	5.83	6.90	0.25	5.7 mg/dL	-0.17	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1778	113.09	4.12	4.88	116 mg/dL	0.64	0.22
SDI Range Interpretation										
Within -1.00 to +1.00 Excellent										
Within ±1.01 to ±2.00 Good										
Within ±2.01 to ±2.99 Accept with caution. Warning Signal										
Beyond ±3.0 Unacceptable performance. Action Signal										

LAB ADDRESS :
 Govt. PHC, Vellalapati
 Vellalapati primary health center
 Vellalapati
 TAMILNADU636912

Ravi Chidambaram
 Dr. Pamela Christy
 QAP COAS Coordinator

for X

05/02, 12:29 PM



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023

Internal Quality Assurance Scheme - First Monthly Summary



Lab Name: Govt. PHC, Vellalappatti
Consistent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No:
Type of Issue Entered:
Date of Report Published:

12003
2023/06/03
2023/06/03

Sl No	Analyte	Method / Principle Name	Reference Range	Accepted Range	Lab Result	CV	CV	CV	CV	CV	CV
1	GLUCOSE III	GOD-POD III	Random (Semi automatic / Fully Automatic) Random (Semi automatic / Fully Automatic)	1823	288.63	5.17	14.82	273 mg/dL	4.82	8.89	0.81
2	CREATININE III	Jaffe Kinetic - Alkaline Picrate III	Random (Semi automatic / Fully Automatic) Random (Semi automatic / Fully Automatic)	1804	2.57	6.64	9.17	2.5 mg/dL	9.41	8.81	0.81
3	CHOLESTROL III	CHOD-PAP III	Random (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.87	106 mg/dL	1.52	8.28	0.28
Interpretation											
Within -1.00 to +1.00 Excellent!											
Within ±1.01 to ±2.00 Good.											
Within ±2.01 to ±2.99 Accept with caution, Warning Signal.											
Beyond ±3.0 Unacceptable performance, Action Signal.											

LAB ADDRESS :
Govt. PHC, Vellalappatti
Vellalappatti primary health center
Vellalappatti
TAMILNADU 625012

Ravula Christa Devi
Dr. Pamela Christoforos
CMC EQAS Coordinator
Christian Medical College, Vellore

Coordinator Contact Details:
Email: cmceqa@cmcvellore.ac.in
Contact Number: 0418-2822102

Dr. Pamela Christoforos
CMC EQAS Coordinator
Christian Medical College, Vellore



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2023



Lab Name: Govt. PHC, Vellalapatti
Constituent Group: Chemistry III
PT Item: Lyophilized Serum

Lab No: 12603
Date of Result Entered: 19/01/2023
Date of Report Published: 09/02/2023

Sl. No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	5.38	7.85	124 mg/dL	0.54	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	1.9 mg/dL	-1.40	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	120 mg/dL	1.15	0.32

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS

Govt. PHC, Vellalapatti
Vellalapatti primary health center
Vellalapatti
TAMILNADU636012

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****