



No.389, Dharsha Garden, Kandappa Colony, Annadanapatty, Salem - 2.  
Cell : 94880 30384, 87787 93052 Email : mallnaveentech@gmail.com Web : www.naveentechnologies.in

CALIBRATION REPORT	
Report No : NT/CAL/23-24/VGR/19	Calibration Date : 28.08.2023
	Calibration valid: 27.02.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, VEERAGANUR, ATTUR, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	HANDY PIPETTE
ii Model	-----
iii Value	100µl-1000µl
iv Serial No	PHC/VGR/MP/002

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.05	1000.13
2	100.03	500.06	1000.11
3	100.04	500.08	1000.09
4	100.08	500.09	1000.07
5	100.01	499.95	1000.05
6	100.09	500.03	1000.01

For Naveen Technologies,

(Authorized signatory)

