

NAVEEN TECHNOLOGIES



No.389, Dharsha Garden, Kandappa Colony, Annadanapatty, Salem - 2.
Cell : 94880 30384, 87787 93052 Email : mailnaveentech@gmail.com Web : www.naveentechnologies.in

CALIBRATION REPORT	
Report No: NT/CAL/23-24/VGR/16	Calibration Date: 28.08.2023
	Calibration Valid: 27.02.2024
I.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, VEERAGANUR, ATTUR, SALEM.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I Manufactured by	HANDY PIPETTE
II Model	-----
III Value	100µl
IV Serial No	PHC/VGR/MP/001

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)	
		100µl
1	100.05	
2	100.06	
3	100.01	
4	99.97	
5	100.01	
6	100.02	

For Naveen Technologies,

S. Ganesh Babu

(Authorized signatory)

