

# NAVEEN TECHNOLOGIES



No.389, Dharsha Garden, Kandappa Colony, Annadanapatty, Salem - 2.  
Cell : 94880 30384, 87787 93052 Email : mailnaveentech@gmail.com Web : www.naveentechnologies.in

CALIBRATION REPORT	
Report No : NT/CAL/23-24/VGR/17	Calibration Date : 28.08.2023 Calibration valid : 27.02.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, VEERAGANUR, ATTUR, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNEPIPETTE
iii Value	10 µl -100 µl
iv Serial No	RW09104


**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Naveen Technologies,

  
(Authorized signatory)

