



CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date : 29.09.2023
	Calibration valid : 28.03.2024
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVERNMENT PRIMARY HEALTH CENTRE,THEVUR , SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I Manufactured by	DRAGON LAB
II Model	5µl - 50µl
III Serial No	YE17AAG0039121
IV Location	LABORATORY

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05µl	20µl	50µl
1	05.01	20.03	50.04
2	05.03	20.04	50.02
3	05.02	20.05	50.09
4	05.03	20.02	50.02
5	05.01	20.06	50.05
6	05.02	20.02	50.04

For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com