



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 29.09.2023 Calibration valid : 28.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, GOVERNMENT PRIMARY HEALTH CENTRE,THEVUR , SALEM.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMO SCIENTIFIC
ii Model	FINEPIPETTE F3
iii Value	10 µl -100 µl
iv Serial No	RW09054

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.04	50.01	100.05
2	10.04	50.05	100.05
3	10.03	50.02	100.04
4	10.02	50.04	100.03
5	10.00	50.02	100.04
6	10.02	50.07	100.02

For Heart Beat Medical Systems,

(Authorized signatory)  

MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com