



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023

External Quality Assurance Scheme  
10812  
18/05/2023  
05/06/2023



Lab Name: Govt. UPHC, Thirunagar  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 10812  
Date of Result Entered: 18/05/2023  
Date of Report Published: 05/06/2023

SI No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	177.3 1.87	-0.38	0.44		
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	117.05 117.05	0.83	0.01		
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83		0.42	0.27		
<b>SDI Range Interpretation</b>												
				Within -1.00 to +1.00							Excellent.	
				Within ±1.01 to ±2.00							Good.	
				Within ±2.01 to ±2.99							Accept with caution. Warning Signal.	
				Beyond ±3.0							Unacceptable performance. Action Signal.	

**LAB ADDRESS :**  
Govt. UPHC, Thirunagar  
MEDICAL OFFICER, THIRUNAGAR  
Madurai-Corp  
TAMILNADU625006

**Coordinator Contact Details:**  
Email: [clingqc@cmcvellore.ac.in](mailto:clingqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
**Dr. Pamela Christudoss**  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components