



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 15.09.2023
	Calibration valid : 14.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVERNMENT PRIMARY HEALTH CENTRE, KOOTTATHUPATTI, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	HANDYPETTE
ii Model	-
iii Value	100µl-1000µl
iv Serial No	120900095

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.05	500.11	1000.09
2	100.04	500.06	1000.04
3	100.02	500.03	1000.07
4	100.06	500.04	1000.14
5	100.02	500.03	1000.11
6	100.05	500.07	1000.04

For Heart Beat Medical Systems,


(Authorized signatory)

