



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 29.09.2023
		Calibration valid : 28.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, GOVERNMENT PRIMARY HEALTH CENTRE,THEVUR , SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	BDI-152
iii	Serial No	PHC/TV/CENT/001

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1001	1
2	1200	1201	1
3	1500	1502	2
4	2000	2003	3

Acceptable deviation is ± 5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com