

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Sriram Nagar	<a href="#">Details About Robust Analysis</a>	
Lab No	10687	<a href="#">Detail About Monthly Summary</a>	
Month	August	<a href="#">Detail about SDI</a>	
Year	2023		
Constituent Group	Chemistry III		
<a href="#">Click on the analyte to view Graphical Data</a>		<a href="#">All Analyser Result</a>	<a href="#">Print</a>
Date of Result Entered : 20/08/2023			
Date of Report Published : 01/09/2023			

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	152 mg/dL	-2.62	0.42
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	117 mg/dL	0.69	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>