

<u>CA</u>	LIBRATIO	N CER	TIFICA	TE		
CERTIFICATE NO: SBS/CL/23/13051		MEDICAL DEVICES		/ICES	Page	No:1 of 1
Issue Date	07-1	0-2023				
SRF No & Date	SRF	23/00659-00	01 & 06-10-2	023		
Receipt Date	06-1	06-10-2023				
Calibration Date	06-1	06-10-2023				
Calibration Due	05-10	05-10-2024				
Customer Name & Address						
GOVERNMENT PRIMARY HEALTH CENT	RE,					
KOOVATHUR-603305, KANCHIPURAM DIS	TRICT.					
	Details of Device	Under Calibra	tion (DUC)			
Description : SEMI AUTO ANALYZER		Make & Model : ROBONIK & PRIETEST TOUCH				
Range : MULTI	Sr. N	Sr. No : ATCD0191220RBK				
Resolution : MULTI	Identi	Identification No		NA		
DUC Condition : SATISFACTORY	Locat	ion	:	LABORATORY		
Enviro	nmental Conditions	& Calibration	Procedure D	etails		
Environmental Details Temperature:25.6 ° C		Relative Humidity 52% RH				
Calibration Procedure No SBS/CP/MD/20		Calibration done at ONSITE				
	Reference	Standards Det	ails			
S.No Description Make/ SI No:		Certificate No		0		alidity
1 Electrical Safety Analyser Rigel Medical		L-1059 M-230809-16-4			0-08-2024	

ELECTRICAL S	SAFETY
---------------------	--------

RESUL	<u>.TS</u>				
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)		
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)		
	>20MΩ	97	13.92		
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)		
	<5000µAfor B,BF,CF	125	8.0		
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)		
	<500µAfor B,BF,CF	235	7.2		

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY





Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.