

CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	
Calibration Date: 28.09.2024	
Calibration Valid: 27.03.2024	
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, URBAN PRIMARY HEALTH CENTRE, ATTUR, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I	Manufactured by
	THERMOSCIENTIFIC
II	Model
	FINNPIPETTE F3
III	Value
	1000µl
IV	Serial No
	RW03147

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	1000µl
1	1000.05
2	1000.06
3	1000.01
4	999.97
5	1000.01
6	1000.02

For Heart Beat Medical Systems,

(Authorized signatory)