

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Sriram Nagar	Details About Robust Analysis								
Lab No	10687	Detail About Monthly Summary								
Month	April	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data										
		<a href="#">All Analyser Result</a>	<a href="#">Print</a>							
Date of Result Entered : 20/04/2023										
Date of Report Published : 06/05/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	115 mg/dL	-0.62	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2 mg/dL	-0.91	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	124 mg/dL	2.14	0.28
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>