



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. PHC, Poolavari Lab No 12626  
 Constituent Group Chemistry III Date of Result Entered : 20/09/2023  
 PT item Lyophilized human serum based Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	308 mg/dL	-0.19	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.28 mg/dL	-0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	115 mg/dL	0.32	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 Govt. PHC, Poolavari  
 Poolavari  
 TAMILNADU636010

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
EQAS YEARLY SUMMARY - 2022



Lab Name Govt. PHC, Poolavari Lab No 12626  
Constituent Group Chemistry III  
PT item Lyophilized human serum based Date of Report Published 20-02-2023

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.88	A
2	CREATININE III	Group I	1927	0.86	A
3	CHOLESTROL III	Group I	1923	1.13	B

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond $\pm 3.00$	D	Unacceptable

#### LAB ADDRESS

Govt. PHC, Poolavari  
Poolavari  
TAMILNADU636010

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College,  
Vellore

Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*