

				CAL	IBRA1	TION	CERT	TIFI(CA	<u>TE</u>	
CERT	IFICATE N	IO: SB	S/CL/23/	13460			M	IEDICA	L DE	VICES	Page No:1 of 1
Issue Date						07-10	-2023				
SRF N	No & Date					SRF/	SRF/23/00753-0003 & 07-10-2023				
Recei	pt Date					07-10	07-10-2023				
Calibr	ration Date	•				07-10	07-10-2023				
Calibr	ration Due					06-10	-2024				
Custo	mer Name	& Addr	ess								
GOVE	ERNMENT	PRIMA	ARY HEA	LTH CENTR	E,						
THUN	/BIVADI-6	39002,	KARUR D	DISTRICT.							
					Details of D	evice U	nder Calibra	tion (D	UC)		
Descri	iption		ELECT	RICAL Y(MICROSCO	OPE)	Make	& Model			ZOOM SCIENTIFIC &	NA
Range		:	MULTI		,	Sr. No				474427	
Resolution		4	MULTI				fication No			NA	
DUC Condition			SATISFACTORY			Location				LABORATORY	
				Enviror	mental Con		& Calibration	n Proce	dure		
Enviro	onmental De	tails		Temperature	25.6°C		Relative Hu	midity		52% RH	
Calibration Procedure No SBS/CP/MD/					Calibration done at			ONSITE			
					Refe	rence S	tandards De	etails			
S.No	S.No Description			Make/ SI N	Make/ SI No:		Certificate No			Validity	
1 Electrical Safety Analyser			Rigel Medic	Rigel Medical & 44L-1059		M-230809-16-4			10-08-202		

ELECTRICAL SAFETY

RESULTS			
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	78	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	204	Officertainty in 76 (I)
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
	<500µAfor B,BF,CF	214	8.9

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

(Calibration Engineer)
K.SATHYAMOORTHY



-	Technical Manager
	C.SHANMUGARAJ

Chief Executive
