

				CALI	BRAT	TION	CER1	TIFICA 7	<u>E</u>	
CERTIFICATE NO: SBS/CL/23/13458					MECHANICAL		Page No:1 of 1			
Issue Date				07-10-2023						
SRF No & Date					SRF/23/00753-0001 & 07-10-2023					
Receipt Date					07-10-2023					
Calibration Date					07-10-2023					
Calibration Due					06-10-2024					
	ner Name &	Addre	ss							
GOVE	RNMENT F	PRIMA	RY HEALTH	CENTRE,						
THUM	BIVADI-63	9002,K	ARUR DISTE	RICT.						
				De	etails of De	evice Und	ler Calibra	ation (DUC)		
Descri	iption : CENTRIFUGE		JGE		Make & Mode		:	B.D INSTRUMENTS & BDI-152		
Range :		3000	RPM		Sr. No		:	NA		
Resolution :		1 RPM			Identification No			NA		
DUC Condition :			Satisfactory			Location			LABORATORY	
	OHOHOH				Condition	s & Stan	dard Ope	rating Proce	dure Details	
Environmental Details Temperature				ire: 24.3 °C			Humidity 54 % Rh			
Calibration Procedure No SBS/CP/ML/				IL/04			done at ONSITE			
					Refer	ence Sta	ndards D	etails		
S.No	Description	scription			Make/ S	Make/ SI No:		Certificate No		Validity
	Digital Tachometer			Lutron 8	Lutron & Q638380		JRPM-CCTR-A&S-2022-0036		03-10-2023	

## CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	1000	999.7	0.3	4.2	
2	2000	1999	1	4.2	
3	3000	2998	2	4.2	

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)

K.SATHYAMOORTHY

Authorised Signatory

Quality Manager

(C.SIVABALAN)

Chief Executive