



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date 12.09.2023
		Calibration valid 1 1.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/S, GOVERNMENT PRIMARY HEALTH CENTRE, THIRUVENKATANATHAPURAM,, THIRUNELVELI
1.2. DESCRIPTION OF DEVICE UNDER TEST (OUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	
ii	Model	BDI 152
iii	Serial No	ZBEN-13706

CALIBRATION WORKSHEET LOCATION: LABORATORY

Tabulation For — Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	
2	1200	1201	1
3	1500	1501	
4	2000	2002	2

Acceptable deviation is ± 5 RPM it depends on

Max Speed Range

D Reg No:
√/CBE/MD42/00016

99400 28355,
Road, Peelamedu,

9751206106No.42/1, Navaindia

GSTIN:
33AAJFH1833MIZB

88833 15662CBE-04,hbmedicalsyst@gmail.com