

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Nasiyanoor Lab No 12197

Constituent Date of Result Entered: 20/09/2023 **Chemistry III** Group

PT item 05/10/2023 Lyophilized human serum based Date of Report Published :

CLN	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	CDI		
SI.No						CV	SD	Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	307 mg/dL	-0.26	0.69	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	118 mg/dL	1.01	0.20	

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. PHC, Nasiyanoor THE MEDICAL OFFICER, ERODE Nasiyanoor **TAMILNADU**

Email:clinqc@cmcvellore.ac.in

Panela Christudos **Coordinator Contact Details:** Dr. Pamela Christudoss **CMC EQAS Coordinator**

Contact Number: 0416-2283102 **Christian Medical College, Vellore**

> Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components ****** End of Report *****