

C	ALIBRA	TION CER	TIFICA?	<u>re</u>	
CERTIFICATE NO: SBS/CL/23/13421	3 (7)	11-11-11-11-11	MEDICAL DEV	/ICES	Page No:1 of 1
Issue Date SRF No & Date	1/2	07-10-2023 SRF/23/00745-0	001 & 07-10-2	2023	
Receipt Date		07-10-2023			
Calibration Date					
Calibration Due		06-10-2024		gelden Alan Santa Comment	
Customer Name & Address					
GOVERNMENT PRIMARY HEALTH C					
ESANATHAM-639203, KARUR DISTRI					
	Details of	Device Under Calibr	ation (DUC)		
Description SEMI AUTO A	NALYZER	Make & Model	:	ROBONIK & PRIET	EST TOUCH
Range MULTI		Sr. No	: 17	AT0500509RBK	
Resolution MULTI		Identification No	1	NA	
DUC Condition SATISFACTO		Location	:	LABORATORY	
E	vironmental Cor	ditions & Calibratio	n Procedure D	Details	
Environmental Details Temperature: 25.6 ° C		Relative Humidity		52% RH	
ibration Procedure No SBS/CP/MD/20		Calibration done at		ONSITE	
Cambradon i Toccadio i to		erence Standards D	etails		
S.No Description	Make/ SI		Certificate N	No	Validity
	fety Analyser Rigel Medical &		M-230809-16-4		

ELECT	RICAL	SAF	ETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20ΜΩ	90	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
-	<5000µAfor B.BF,CF	184	7.5
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
	SUOUAfor B.BF.CF	226	7.2

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6 Equipment used for Calibration were calibrated & traceable to National & International Standards. Calibrated By

**Authorised Signatory** 

(Calibration Engineer) K.SATHYAMOORTHY





Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.