



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13488	MEDICAL DEVICES	Page No:1 of 1
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00686-0003 & 06-10-2023	
Receipt Date	06-10-2023	
Calibration Date	06-10-2023	
Calibration Due	05-10-2024	

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE(SIRUVANTHADU BLOCK),
P.S.PALAYAM-605502,VILLUPURAM DISTRICT.

Details of Device Under Calibration (DUC)

Description	: SEMI AUTO ANALYZER	Make & Model	: ROBONIK & PRIETEST TOUCH
Range	: MULTI	Sr. No	: ATCD2310321RBK
Resolution	: MULTI	Identification No	: NA
DUC Condition	: SATISFACTORY	Location	: LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 89	Uncertainty in % (±) 13.92
2	Earth Leakage <5000μAfor B,BF,CF	Measured values in μA 146	Uncertainty in % (±) 7.9
3	Enclosure Leakage <500μAfor B,BF,CF	Measured values in μA 223	Uncertainty in % (±) 7.3

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory


(Calibration Engineer)
M.DINESH




 Technical Manager
C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/13489	MECHANICAL	Page No:1 of 1
Issue Date	07-10-2023		
SRF No & Date	SRF/23/00686-0004 & 06-10-2023		
Receipt Date	06-10-2023		
Calibration Date	06-10-2023		
Calibration Due	05-10-2024		

Customer Name & Address
 GOVERNMENT PRIMARY HEALTH CENTRE(SIRUVANTHADU BLOCK),
 P.S.PALAYAM-605502,VILLUPURAM DISTRICT.


Details of Device Under Calibration (DUC)			
Description	CENTRIFUGE	Make & Model	LABOTECH & BDI-152
Range	3500 RPM	Sr. No	NA
Resolution	1 RPM	Identification No	NA
DUC Condition	Satisfactory	Location	LABORATORY

Environmental Conditions & Standard Operating Procedure Details			
Environmental Details	Temperature: 25.4°C	Relative Humidity	54% Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024


CALIBRATION RESULTS				
S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	998.6	1.4	4.2
2	2000	1998.6	1.4	4.2
3	3000	2998.7	1.3	4.2

- REMARKS**
- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - 2.The user should determine the suitability of the instrument for its intended use.
 - 3.The recalibration interval should be determined on the user requirement.
 - 4.The results stated in this certificate relate only to the item calibrated.
 - 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

 (Calibration Engineer)
 M.DINESH



Authorised Signatory

 Quality Manager
 (C.SIVABALAN)
 Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13490		MEDICAL DEVICES		Page No:1 of 1
Issue Date	07-10-2023			
SRF No & Date	SRF/23/00686-0005 & 06-10-2023			
Receipt Date	06-10-2023			
Calibration Date	06-10-2023			
Calibration Due	05-10-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE(SIRUVANTHADU BLOCK), P.S.PALAYAM-605502,VILLUPURAM DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	ELECTRICAL SAFETY(MICROSCOPE)	Make & Model	LABOMED & CXL PLUS	
Range	MULTI	Sr. No	130661113	
Resolution	MULTI	Identification No	NA	
DUC Condition	SATISFACTORY	Location	LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.5°C	Relative Humidity	55% RH	
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 90	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 175	Uncertainty in % (±) 7.5
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 198	Uncertainty in % (±) 7.3

REMARKS


- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory


(Calibration Engineer)
M.DINESH




Technical Manager
C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.