

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13166

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE(SIRUVANTHADU BLOCK), P.S.PALAYAM-605502,VILLUPURAM DISTRICT.	SRF No.	SRF/23/00686-0001
	SRF Date	06-10-2023
	Date of Receipt	05-10-2023
	Date of Calibration	06-10-2023
	Due Date for Calibration	05-10-2024
	Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	MICROLUX
Range	5-50µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	5-50µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP/22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
		1	10	9.87	9.86	9.87				
		9.88	9.86	9.87	9.88	9.89				
2	30	29.76	29.77	29.78	29.79	29.76	29.78	-0.22	0.01	0.47
		29.77	29.78	29.79	29.78	29.78				
3	50	49.88	49.87	49.88	49.87	49.88	49.87	-0.13	0.01	0.47
		49.87	49.86	49.86	49.87	49.88				


Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grades as specified in ISO 3696.

Calibrated By,

M. Adhiban
(Calibration Engineer)

M.ADHIBAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13167

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE(SIRUVANTHADU BLOCK), P.S.PALAYAM-605502,VILLUPURAM DISTRICT.	SRF No.	SRF/23/00686-0002
	SRF Date	06-10-2023
	Date of Receipt	05-10-2023
	Date of Calibration	06-10-2023
	Due Date for Calibration	05-10-2024
	Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	MICROLUX
Range	100-1000µl	Model	NA
Resolution	10µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP12/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	100	99.92	99.94	99.91	99.90	99.92	99.92	-0.08	0.01	0.47	
		99.94	99.93	99.92	99.91	99.92					
2	500	499.04	498.99	499.05	499.02	499.04	499.01	-0.99	0.03	0.47	
		498.96	498.99	499.02	499.03	498.99					
3	1000	998.99	998.89	998.90	998.91	998.91	998.93	-1.07	0.03	0.47	
		998.96	998.95	998.92	998.94	998.93					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade 3 as specified in ISO 3696.

Calibrated By,

M. Adhiban
(Calibration Engineer)
M.ADHIBAN



Authorised by:

C. Sivabalan
Quality Manager/Chief Executive
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.