

**CALIBRATION CERTIFICATE**

Certificate No: SBS/CL/23/13483

Page No : 1 of 1

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE, SIRUKUNDRUM-603108,CHENGALPATTU DISTRICT.	SRF No.	SRF/23/00760-0001
	SRF Date	06-10-2023
	Date of Receipt	06-10-2023
	Date of Calibration	06-10-2023
	Due Date for Calibration	05-10-2024
	Issue Date	07-10-2023

**Details of Unit Under Calibration**

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	QW10063	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

**Environmental Condition**

**Calibration Method Used**

Temperature (°C)	23.6	Humidity (%RH)	54	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

**Result of Calibration in µl**


Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10.0	9.90	9.92	9.88	9.91	9.92	9.91	-0.09	0.01	0.47
		9.91	9.92	9.93	9.92	9.90				
2	50.0	49.78	49.76	49.74	49.72	49.74	49.74	-0.26	0.02	0.47
		49.76	49.75	49.73	49.72	49.71				
3	100.0	98.92	98.94	98.94	98.96	98.92	98.93	-1.07	0.02	0.47
		98.91	98.93	98.94	98.92	98.91				

**Remarks**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

*M. Adhiban*  
(Calibration Engineer)  
M.ADHIBAN



Authorised by:

*C. Sivabalan*  
(Quality Manager/Chief Executive)  
C.SIVABALAN

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.