

CALIBRATION REPORT

Report No: HBMS/CAL/23-24		Calibration Date: 04.10.2023
I.1. CUSTOMER DETAILS:		Calibration valid: 03.04.2024
Name and Address of the organization		M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE, PONNAMAPATTI, SALEM CORPORATION, SALEM.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		
i	Manufactured by	CENTRIFUGE
ii	Model	REMI
iii	Serial No	R-8C
		ZEAN-03148

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For - Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	0
2	1200	1202	2
3	1500	1501	1
4	2000	2003	3

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com