




Trust & Care



heartbeat

Medical Systems

CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 04.10.2023
		Calibration valid : 03.04.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE , PONNAMAPATTI, SALEM CORPORATION, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMO SCIENTIFIC
ii	Model	FINE PIPETE F3
iii	Value	100µl-1000µl
iv	Serial No	RW11530


Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.03	1000.03
2	100.01	500.01	1000.02
3	100.05	500.07	1000.02
4	100.03	500.05	1000.08
5	100.03	500.05	1000.09
6	100.05	500.07	1000.04

For Heart Beat Medical Systems,



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com