

HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),
SIVAKASI, VIRUDHUNAGAR-626124.
E-mail : hopeultrasystems@gmail.com

GSTIN. 33DLMPK3057D1Z9
Mobile: +919487402761, +919715223192.

CALIBRATION CERTIFICATE

Requested By

GPHC, VANNIAMPATTI.

Scope:

Equipment Name	: Micropipette	Asset no	: PHC/VP-LAB/002
Make	: Microlux	Calibration Done at	: customer place
Model	: 100-1000 μ l	Calibration Date	: 28/08/2023
Sl no	: ...	Next call due on	: 27/08/2024

TEST RESULTS Volume μ l

Sl. No.	DUC reading(μ l)	Standard Reading (μ l)		Correction (μ l)
		Trail1	Trail 2	
01	100	100.0	100.0	+0
02	200	200.0	200.0	+0
03	500	500.0	500.0	+0
04	800	800.0	800.0	+0
05	1000	1000.0	1000.0	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

1. Electronic weighing scale(EWC/SF400C)

Note:

1. Calibration points were selected as per customer request.
2. Each reading was taken by repeatability.
3. Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems

Authorized signatory

