CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023

Lab Name	Govt. PHC, Endiyur	Lab No	13004
Consultant Group	Chemistry III	Date of Result Entered :	19/08/2023
PT Item	Lyophilized human serum based	Date of Report Published	01/09/2023

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1641	175.58	5.13	9.09	175 mg/dL	4.55	0.42
2	CREATININE III	Jaffe Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	0.77	0.12	1.8 mg/dL	1.19	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.87	4.64	121 mg/dL	1.55	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.99	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS
Govt. PHC, Endiyur
Endiyur
TAMILNADU

Coordinator Contact Details:
Email:ctmqc@cmcvellore.ac.in
Contact Number: 0416-2283192

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
****** End of Report ******

Dr. K. DIVYA
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
ENDIYUR
REG. NO. 105905