



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE
SCHEME

EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name **Govt. PHC, Sathyanagar** Lab No **12587**
 Constituent Group **Chemistry III**
 PT item **Lyophilized human serum based** Date of Report Published **20-02-2023**

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.53	A
2	CREATININE III	Group I	1927	1.50	B
3	CHOLESTROL III	Group I	1923	0.90	A

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond ± 3.00	D	Unacceptable

LAB ADDRESS

Govt. PHC, Sathyanagar
 SATHYA NAGAR PHC LAB
 Sathyanagar
 TAMILNADU636303

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact number:0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College,
 Vellore

Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****