

# MEP Automation

55, West Pradakshnam Road, Karur - 639 001.

## Calibration Certificate for Semiautoanalyser

Customer : BMO , Innungur UPHC , Karur dist

Calib. Date : 04.01.2023

Due date : 03.01.2024

Model : Robonik touch

Serial no : ATCD396032RBK

### Calibraiotn Standrads Used

Name	Model	Srno
Digital Multimeter:	MASTECI	M-830 BZ

### I .Measuring Reports

Test	Limit	Result
Supply Voltage (Line -Neutral)	210 - 240 V	230 V
Supply Voltage (Line -Ground )	210 - 240 V	235 V
Supply Voltage (Neutral- Ground)	< 5V	2 V

I. Lamp module checked ok

II. Asipitraion motor head checked ok

Tested and Calibrated by



# MEP Automation

55, West Pradakshnam Road, Karur - 639 001.

## Calibration Certificate for Centrifuge

Customer : BMO , Innungur UPHC , Karur dist

Calib. Date : 04.01.2023

Due date : 03.01.2024

Model : Electric

Serial no : cent / 001

### Calibraiotn Standrads Used

Name	Model	Srno
Digital Multimeter:	MASTECI	M-830 BZ

### I .Measuring Reports

Test	Limit	Result
Supply Voltage (Line -Neutral)	210 - 240 V	230 V
Supply Voltage (Line -Ground )	210 - 240 V	235 V
Supply Voltage (Neutral- Ground)	< 5V	2 V

I. Carbon brushes checked ok

II. Rotar head checked ok

Tested and Calibrated by

