

CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/13335	MECHANICAL	Page No:1 of 1
Issue Date	07-10-2023		
SRF No & Date	SRF/23/00725-0002 & 07-10-2023		
Receipt Date	07-10-2023		
Calibration Date	07-10-2023		
Calibration Due	06-10-2024		

Customer Name & Address
 GOVERNMENT PRIMARY HEALTH CENTRE,
 THDGAMALAI,KAVALKARANPATTI-621313,KARUR DISTRICT,


Details of Device Under Calibration (DUC)			
Description	CENTRIFUGE	Make & Model	LINCO & NA
Range	3500 RPM	Sr. No	CN-35/385
Resolution	1 RPM	Identification No	NA
DUC Condition	Satisfactory	Location	LABORATORY

Environmental Conditions & Standard Operating Procedure Details			
Environmental Details	Temperature: 25.3°C	Relative Humidity	55% Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

Reference Standards Details				
S No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024


CALIBRATION RESULTS				
S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	998.8	1.2	4.2
2	2000	1999.3	0.7	4.2
3	3500	3499.4	0.6	4.2

- REMARKS**
- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - The user should determine the suitability of the instrument for its intended use.
 - The recalibration interval should be determined on the user requirement.
 - The results stated in this certificate relate only to the item calibrated.
 - The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
 - Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

 (Calibration Engineer)
 K SATHYAMOORTHY



Authorised Signatory

 <input type="checkbox"/> (Quality Manager) (C SIVABALAN)	<input type="checkbox"/> (Chief Executive)
---	--

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.