



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. PHC, Chemmani Chettlipyam Lab No 9251
Constituent Group Chemistry III Date of Result Entered : 19/07/2023
PT item Lyophilized human serum based Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	24	109.83	6.83	7.50	115 mg/dL	0.69	3.06
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	18	5.79	3.00	0.17	5.5 mg/dL	-1.67	0.08
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	112.64	4.38	4.93	108 mg/dL	-0.94	2.26

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Chemmani Chettlipyam

THE MEDICAL OFFICER, GOVT PRIMARY HEALTH CENTER, CHEMMANI CHETTIPALAYAM, MGC PALAYAM POST, NEAR GANESHAPURAM, COIMBATORE

Chemmani Chettlipyam

TAMILNADU641107

Pamela Christudoss

Coordinator Contact Details:
Email: cllnqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****