

				CALIB	RAT	101	N CERT	TIFIC.	<u>ATE</u>		
CERTIFICATE NO: SBS/CL/			SBS/CL/2	:3/12406		MECHANICAL		ICAL	Page	No:1 of 1	
Issue Date				05-10-2023							
SRF No & Date				SRF/23/00581-0002 & 04-10-2023							
Receipt Date					04-10-2023						
Calibration Date					04-10-2023						
Calibration Due					03-10-2024						
Custor	ner Name &	Addres	SS_								
GOVE	RNMENT P	RIMAR	Y HEALTH	CENTRE,							
MELVI	HARAM,RA	NIPET	DISTRICT.								
				Detai	Is of De	vice l	Under Calibra	ation (DU	C)		
Description		:	CENTRIFUGE		Make & Model		:	LABOTEC	H & BDI-152		
Range		:	3500 RPM		Sr. No		:	NA			
Resolution :		:	1 RPM			Identification No :		NA			
DUC Condition : Satisf		Satisfactor	ory		Location		:	LABORAT	ORY		
			Envir	onmental Co	nditions	s & S	tandard Ope	rating Pro	cedure Details		
Environmental Details Temperature					e: 25.5°C		Relative Humidity		52%Rh		
Calibration Procedure No SBS/CP/				SBS/CP/ML/	_/04		Calibration done at		ONSITE		
					Refere	ence s	Standards De	etails	•		
S.No	lo Description				Make/ SI No:			Certificate No			Validity
1	Digital Tachometer			LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013			09-06-2024		

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY		
	RPM	RPM	RPM	%		
1	500	499.0	1.0	4.2		
2	1500	1499.5	0.5	4.2		
3	3000	2999.5	0.5	4.2		

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By
(Calibration Engineer)
E.ESWAR



Authorised Signatory

(Quality Manager) (C.SIVABALAN)

(Chief Executive)

SUNSHINE BIOMEDICAL SOLUTIONS