

HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),
SIVAKASI, VIRUDHUNAGAR-626124.

E-mail : hopeultrasystems@gmail.com

GSTIN. 33DLMPK3057D1Z9
Mobile: +919487402761, +919715223192.

CALIBRATION CERTIFICATE

Requested By

GPNC,NATHAMPATTI.

Scope:

Equipment Name : Micropipette
Make : MICROLUX
Model : 10-100 μ l
Sl no : ...

Asset no : PHC/NPT-LAB/003
Calibration Done at : customer place
Calibration Date : 30/08/2023
Next call due on : 29/08/2024

TEST RESULTS Volume μ l

Sl. No.	DUC reading(μ l)	Standard Reading (μ l)		Correction (μ l)
		Trail1	Trail 2	
01	10	10.0	10.0	+0
02	20	20.0	20.0	+0
03	50	50.0	50.0	+0
04	80	80.0	80.0	+0
05	100	100.0	100.0	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

2. Electronic weighing scale(EWC/SF400C)

Note:

4. Calibration points were selected as per customer request.
5. Each reading was taken by repeatability.
6. Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems

Authorized signatory

