

HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),
SIVAKASI, VIRUDHUNAGAR-626124.

E-mail : hopeultrasystems@gmail.com

GSTIN. 33DLMPK3057D1Z9

Mobile: +919487402761, +919715223192.

CALIBRATION CERTIFICATE

Requested By

GPHC,NATHAMPATTI.

Scope:

Equipment Name : CENTRIFUGE
Make : ...
Model : ...
Sl no :

Asset no : PHC/NPT-LAB/001
Calibration Done at : customer place
Calibration Date : 30/08/2023
Next call due on : 29/08/2024

<u>Sl. No.</u>	<u>DUC reading(RPM)</u>	<u>Standard Reading (RPM)</u>	<u>Correction (RPM)</u>
01	Position 1	2010	+0
02	Position 2	2140	+0
03	Position 3	2460	+0
04	Position 4	3250	+0
05	Position 5	3500	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

1. Digital Multimeter-Mastech-MS830L
2. Digital Taco meter-DT-2234C/Max tech

Note:

1. Calibration points were selected as per customer request.
2. Each reading was taken by repeatability.
3. Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems



Authorized signatory