




Trust & Care



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 03.10.2023 Calibration valid : 02.04.2024
I.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE, KITCHIPALAYAM, SALEM CORPORATION, SALEM.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	HANDYPETTE
ii Model	--
iii Value	10 µl -100 µl
iv Serial No	121701149


Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report

Number of measurement	Channel I (Volume)		
	10µl	50µl	100µl
1	10.02	50.02	100.05
2	10.02	50.04	100.02
3	10.01	50.02	100.02
4	10.02	50.01	100.05
5	10.00	50.02	100.04
6	10.02	50.03	100.02

For Heart Beat Medical Systems,


 (Authorized signatory)

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com

CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 03.10.2023 Calibration valid : 02.04.2024
I.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE , KITCHIPALAYAM, SALEM CORPORATION, SALEM.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	HANDYPETTE
ii Model	-
iii Value	100µl-1000µl
iv Serial No	11802202


Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.02	1000.02
2	100.02	500.01	1000.04
3	100.04	500.05	1000.07
4	100.02	500.05	1000.08
5	100.02	500.05	1000.05
6	100.05	500.02	1000.04

For Heart Beat Medical Systems,



MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com