

			CALIE	BRAT	10	N CER	TIFI	CAT	TE_		
CERTIFICATE NO: SBS/CL/23/13			BS/CL/23/13515	13515		MECHANICAL			L	Page	No:1 of 1
Issue Date				06-10-2023							
SRF No & Date				SRF/23/00672-0005 & 06-10-2023							
Receipt Date				06-10-203							
Calibration Date				06-10-2023							
Calibration Due					05-10-2024						
_	mer Name & A										
GOVE	RNMENT UR	BAN PRI	MARY HEALTH CEN	NTRE,			2.47				
KEEZI	HPERUMPAK	KAM-605	602,VILLUPURAM [DISTRICT.							
			Deta	ils of Dev	ice l	Under Calibr	ration (DUC)			
Descri	ption	: CI	ENTRIFUGE		Make	e & Model		:	REMI & R-8C		
Range		:	5250 RPM		Sr. N	lo		:	ZEBN-04619		
Resolu	ution	:	10 RPM		ldent	ification No		:	NA		
DUC C	Condition	: Sa	atisfactory		Loca			:	LABORATOR	RY	
			Environmental C	onditions	& S1	andard Ope	rating	Proced	ure Details		
Environmental Details Temperature			e: 25.5°C Relative H		Relative Hur	midity 52%Rh		52%Rh			
Calibration Procedure No SBS/CP/			SBS/CP/ML	_/04 Calib		Calibration d	alibration done at		ONSITE		
				Referer	ice S	Standards De	etails		•		
S.No	S.No Description			Make/ SI No:		Certificate No				Validity	
1	Digital Tachometer			LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013			3	09-06-2024	

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	1000	999.0	1.0	4.2	
2	2000	1999.5	0.5	4.2	
3	3500	3499.5	0.5	4.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY



Authorised Signa	itory
Quality Manager)	(Chief F

(C.SIVABALAN)

(Chief Executive)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.