



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13221	MEDICAL DEVICES	Page No:1 of 1
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00698-0001 & 06-10-2023	
Receipt Date	06-10-2023	
Calibration Date	06-10-2023	
Calibration Due	05-10-2024	

Customer Name & Address
GOVERNMENT PRIMARY HEALTH CENTRE,
SANKARANPANDAL-609308,MAYILADUTHURAI DISTRICT.

Details of Device Under Calibration (DUC)			
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH
Range	MULTI	Sr. No	ATCD4281220RBK
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 80	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 164	Uncertainty in % (±) 7.6
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 216	Uncertainty in % (±) 7.3

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

K. Sathyamoorthy
(Calibration Engineer)
K.SATHYAMOORTHY



Authorised Signatory

C. Shanmugaraaj
 Technical Manager
C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/13222	MECHANICAL	Page No:1 of 1
Issue Date	07-10-2023		
SRF No & Date	SRF/23/00698-0002 & 06-10-2023		
Receipt Date	06-10-2023		
Calibration Date	06-10-2023		
Calibration Due	05-10-2024		

Customer Name & Address
 GOVERNMENT PRIMARY HEALTH CENTRE,
 SAMKARANPANDAL-609308,MAYILADUTHURAI DISTRICT.

Details of Device Under Calibration (DUC)

Description : CENTRIFUGE	Make & Model : LABTECH & NA
Range : 3500 RPM	Sr. No : 76518
Resolution : 1 RPM	Identification No : NA
DUC Condition : Satisfactory	Location : LABORATORY

Environmental Conditions & Standard Operating Procedure Details

Environmental Details	Temperature: 25.5°C	Relative Humidity	54% Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	998.6	1.4	4.2
2	2000	1998.5	1.5	4.2
3	3000	2998.5	1.5	4.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

K.SA
 (Calibration Engineer)
 K.SATHYAMOORTHY



Authorised Signatory

C.Sivabalan
 Quality Manager Chief Executive
 (C.SIVABALAN)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13223

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, SANKARANPANDAL-609308,MAYILADUTHURAI DISTRICT.	SRF No.	SRF/23/00698-0003
	SRF Date	06-10-2023
	Date of Receipt	05-10-2023
	Date of Calibration	06-10-2023
	Due Date for Calibration	05-10-2024
	Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW12505	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP122/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.92	99.94	99.91	99.90	99.92	99.92	-0.08	0.01	0.47
		99.94	99.93	99.92	99.91	99.92				
2	500	499.04	498.99	499.05	499.02	499.04	499.01	-0.99	0.03	0.47
		498.96	498.99	499.02	499.03	498.99				
3	1000	998.99	998.89	998.90	998.91	998.91	998.93	-1.07	0.03	0.47
		998.96	998.95	998.92	998.94	998.93				


Remarks

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming to Grade3 as specified in ISO 3696.

Calibrated By,

M.P. Adhiban
(Calibration Engineer)

M.ADHIBAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13224

Page.No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, SANKARANPANDAL-609308,MAYILADUTHURAI DISTRICT.	SRF No.	SRF/23/00698-0004
	SRF Date	06-10-2023
	Date of Receipt	05-10-2023
	Date of Calibration	06-10-2023
	Due Date for Calibration	05-10-2024
	Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW09769	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10.0	9.94	9.93	9.94	9.92	9.94	9.93	-0.07	0.01	0.47	
		9.92	9.92	9.94	9.94	9.94					
2	50.0	49.87	49.88	49.87	49.88	49.87	49.87	-0.13	0.01	0.47	
		49.88	49.86	49.87	49.86	49.87					
3	100.0	99.86	99.87	99.88	99.87	99.88	99.87	-0.13	0.01	0.47	
		99.87	99.86	99.86	99.88	99.86					

Remarks


- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,


(Calibration Engineer)
M.ADHIBAN



Authorised by:


(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.