



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



Lab Name Govt. Block PHC, Chinnadharapuram Lab No 7410  
 Constituent Group Chemistry III Date of Result Entered : 19/09/2023  
 PT Item Lyophilized human serum based Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	298 mg/dL	-0.86	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	113 mg/dL	-0.14	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. Block PHC, Chinnadharapuram  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, KARUR DISTRICT.  
 Chinnadharapuram  
 TAMILNADU639202

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*