

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. PHC, Arasamangalam Lab No 13072

Constituent Group Chemistry III Date of Result Entered : 20/09/2023

PT item Lyophilized human serum based Date of Report Published : 05/10/2023

CLNa	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	CDI	
SI.No						CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	317 mg/dL	0.41	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	123 mg/dL	2.17	0.20

SDI Range	Interpretation		
Within -1.00 to +1.00	Excellent.		
Within ±1.01 to ±2.00	Good.		
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.		
Beyond ±3.0	Unacceptable performance. Action Signal.		

LAB ADDRESS:

Govt. PHC, Arasamangalam

Arasamangalam TAMILNADU

Panela Christudoss

Coordinator Contact Details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

********* End of Report *******