



## CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/13469	MECHANICAL	Page No:1 of 1	
Issue Date	07-10-2023			
SRF No & Date	SRF/23/00756-0001 & 07-10-2023			
Receipt Date	07-10-2023			
Calibration Date	07-10-2023			
Calibration Due	06-10-2024			
<b>Customer Name &amp; Address</b>				
GOVERNMENT PRIMARY HEALTH CENTRE, MANDAPAM, RAMANATHAPURAM DISTRICT.				
<b>Details of Device Under Calibration (DUC)</b>				
Description	: CENTRIFUGE	Make & Model	: B.D INSTRUMENTATION & BDI-152	
Range	: 3500 RPM	Sr. No	: NA	
Resolution	: 1 RPM	Identification No	: NA	
DUC Condition	: Satisfactory	Location	: LABORATORY	
<b>Environmental Conditions &amp; Standard Operating Procedure Details</b>				
Environmental Details	Temperature: 24.3 °C	Relative Humidity	54 % Rh	
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE	
<b>Reference Standards Details</b>				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	Lutron & DT-2238	JRPM-CCTR-A&S-2022-0036	03-10-2023

### **CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY ( ± )
	RPM	RPM	RPM	%
1	1000	999.9	0.1	4.2
2	2000	1999.8	0.2	4.2
3	3000	2999.7	0.3	4.2

### **REMARKS**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
(Calibration Engineer)  
P.PRASANNA



Authorised Signatory

 Quality Manager  Chief Executive  
(C.SIVABALAN)

## **SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.