



CHRISTIJAN MEDICAL COLLEGE
VELLORE
In Service of the Nation Since 1900...



NEW LAB MONTHLY SUMMARY

Lab Name: Govt. P.H.C., Thogalpet
 Lab No: 52873
 Month: May
 Year: 2023
 Consultant Group: Chemistry III

Click on the analyte to view Graphical Data

Details About Instrument Analysis
 Detail About Monthly Summary
 Detail about SDK

 Date of Result Entered : 20/05/2023
 Date of Report Published : 06/06/2023

SL No	Analyte	Method / Principle Name	Analyser Name	Unit	Mean	SD	Performance		Total Value	SDS
							CV	SD		
1	GLUCOSE III	GCD 900 III	Robotic (Semi automatic / Fully Automatic)	mg/dL	1834	180.85	5.19	9.39	125.1 mg/dL	-0.54
2	CREATININE III	Jaffe's Kinetic - Alkaline Picrate III	Robonic (Semi automatic / Fully Automatic)	mg/dL	1834	1.76	7.46	0.12	1.61 mg/dL	-0.78
3	CHOLESTROL III	CHOD PAP III	Robonic (Semi automatic / Fully Automatic)	mg/dL	1716	114.60	5.08	6.83	113 mg/dL	-0.27
SCL Range Interpretation: Within -1.00 to -1.00 Excellent Within ±1.01 to ±2.00 Good Within ±2.01 to ±2.99 Accept with caution Warning Signal Beyond ±3.0 Unacceptable performance Action Signal										

Prasad
 Medical Officer
 Govt. P.H.C., Thogalpet
 Kandyangal P.O.,
 Thogalpet, Dist. Pinar: 606 304.